

**CORRIGENDUM TO PREBID MEETING HELD ON DT.14.11.2017 REGARDING TENDER
FOR SUPPLY & INSTALLATION OF LABORATORY ITEMS FOR GOVT. MEDICAL
COLLEGE, BALASORE**

Bid ref. No.: 34/17-18/GMCH(BIs) Date: 23.10.2017

A. Queries on Commercial Parameters:

Sl. No.	Queries raised by prospective Bidders	Response to the Queries
1.	As EMD has to be furnished in BG form and most of the nationalised banks are not available in Balasore, the tender submission last date may be extended for another 10 days.	Extended as follows – 1. Last date & time of tender submission – 27.12.2017 (up to 5PM) 2. Date & time of technical bid opening– 3.1.2018 (12 Noon)
2.	In price bid, the VAT/CST may be amended to GST.	Amended. Pl. refer to Commercial Amendment
3.	To consider the award criteria in lot 2 on the basis of individual item as in case of LOT 1.	No change.
4.	Bidder must submit proof of supply for available service centre in Odisha in last one year.	Pl. See Para 6 of Eligibility Criteria. No change.
5.	Demonstration of microscope should be essential under technical bid evaluation criteria.	Pl. See Para 27.3 of General Conditions of Contract. No change.
6.	The last date of tender may be extended by another 10 days.	Extended as follows – 1. Last date & time of tender submission – 27.12.2017 (up to 5PM) 2. Date & time of technical bid opening– 3.1.2018 (12 Noon)

Amendments

The **below mentioned items along with all categories of weighing scales (Micro, Small, Medium & Large)** asked under the various lots/ packages of the tender document under specified Departments are deleted. Hence the bidders are advised not to quote the respective items.

ANATOMY	
SL No.	Item Name
1	Sledge large cutting microtome
2	Weighing Balance (6kg)
PHYSIOLOGY	
1	Electric Steriliser
2	Haemoglobinometer (Sahli's)
3	Simple Spirometer/ Student Spirometer
BIOCHEMISTRY	
1	Weighing balance Semi-micro (0.0001gram to 100gram.)
2	Vortex mixer
3	Urinometer
4	Magnetic stirrer

5	Spectrophotometer
6	Digital Calorimeter
7	Laminar Flow Bench
8	Thermal Cycler Gradient
MICROBIOLOGY	
1	Refrigerated centrifuge
2	Electrophoresis set
3	BOD incubator
4	Laminar flow table
5	VDRL shaker
6	Vertex Mixer
7	Thermometer (Assorted)
COMMUNITY MEDICINE	
1	Ice lined refrigerator
2	Electric steriliser
COMMON ITEMS (LOT-1)	
1	Monocular Microscope
2	Binocular Microscope
3	Semi automatic Rotary Microtome
4	Incubator
5	Hot plate
6	Analytical Weighing Balance
7	Water still for distilled water
8	Hot air oven
9	Stethoscope
10	BP Apparatus
11	Water bath
12	Autoclave
13	Water distillation plant (Glass)
14	Centrifuge Machine
15	Elisa Reader and Washer
16	Deep freezer
17	Paraffin Embedding Bath

1. Technical Parameter:

Queries & subsequent amendment on Technical specification

Lot 1: Common Items

1. Item 20: Automatic Tissue Processor

- a) **Queries:** System may have in built centrifugal or vacuum control.
Amendment: System should have centrifugal or vacuum control.
- b) **Queries:** Capacity of more than 100 cassettes may be specified.
Amendment: Should have capacity of 120 to 150 cassettes or more.
- c) **Queries:** External UPS may not be required in case of inbuilt power back up.
Amendment: In case of availability of inbuilt battery, external UPS need not to be supplied.
- d) **Queries:** Product certification like CE may be asked.
Amendment: Product should be CE certified.

2. Item 21: 5 Part Diff. Random Access Haematology Analyser

- a) **Queries:** Throughput of 80 samples per hour may not be available.
Amendment: Should have throughput of 60 samples per hour.
- b) **Queries:** Minimum sample aspiration volume should be specified.

Amendment: Minimum sample aspiration volume should be atleast 20uL.

- c) **Queries:** Instrument may have fully automated fluorescence flow cytometry based 6-part differential haematology Analyzer. (IG reportable parameters)

Amendment: Instrument should be fully automated fluorescence flow cytometry based 6-part differential haematology Analyzer having IG as reportable parameters.

3. Item 26: Automatic Slide Stainer

- a) **Queries:** May have 8 program of up to atleast 12 steps and 20 specimen slides per rack instead of 15 programmes and 25 steps.

Amendment: Should have atleast 8 programs of up to atleast 12 steps and 20 specimen slides per rack.

- b) **Queries:** Reagent capacity of minimum 450ml.should be there.

Amendment: Should have reagent capacity of minimum 300ml.

- c) **Queries:** Product certification like CE may be asked.

Amendment: Product should be CE certified.

Lot 2/ Package 2: DEPARTMENT OF PHYSIOLOGY

Item Name: Haemocytometer

Queries: Attached suction bulb needs to be deleted.

Amendment: Attached suction bulb is deleted.

Item Name: Large extension kymograph

Queries: Specification may be amended.

Amendment: The revised specification is as follows,

Should have super quality metal gear with cut ratio should be 0.625mm/sec to 40mm/sec. Should be sprig loaded device, manually operatable, rigid cast iron stand. Should have neutralisation clutch for immediate start & stop, centralised electrical points for respiratory pump, time clock and operation table, mounted on heavy metal platform with manometer.

Lot 2/ Package 3: DEPARTMENT OF BIOCHEMISTRY

1. Item Name: Auto analyser (300 throughput)

- a) **Queries:** Throughput should be clearly specified.

Amendment: As mentioned in the item name, throughput should be 300 tests/hr. excluding ISE

- b) **Queries:** Sample position or no. of test performed simultaneously may be reduced from 100 tests.

Amendment: Sample position or no. of test performed simultaneously should be 50 tests or more.

- c) **Queries:** Reagent position may be reduced from 70 or more.

Amendment: Reagent position should be 50 or more.

- d) **Queries:** Sample probe and reagent probe may be different to avoid cross contamination.

Amendment: Sample probe and reagent probe should be separate.

2. Item Name: Auto analyser (150 throughput)

- a) **Queries:** May have onboard laundry system.

Amendment: Should have onboard laundry system for complete cleaning and drying.

- b) **Queries:** Throughput may be higher upto 200 per hour.

Amendment: No Change

Lot 2/ Package 4: DEPARTMENT OF COMMUNITY MEDICINE

Item Name: Dissecting Microscope

The technical specification for this item is deleted and has been amended as follows:

- Should be a monocular microscope.

- Objective Lenses: DIN 2x objective and should be achromatic, parfocal, parcentered, color coded.
- 45° inclined viewing head should rotate 360° for easy repositioning
- Eyepieces: 10x eyepiece.
- Total Magnification: Fixed to 20X
- Specimen Stage: Locked-on spring activated stage clips.
- Illumination: LED illumination with on/off switch
- Should run on rechargeable batteries (included) and should include recharger for 230V power supply.

Lot 2/ Package 5: DEPARTMENT OF PATHOLOGY

1. Item Name: Freezing Microtome

- a) **Queries:** Cryo chamber temp. and specimen cooling temp. range may be upto -35degC.
Amendment: Cryo chamber temp. setting should be of 0 to -35degC and specimen cooling temp. range should be -10 to -35degC.
- b) **Queries:** Microtome section thickness setting range may be 1 to 30um instead of 0.5 to 30um.
Amendment: Microtome section thickness setting range may be 1 to 30um.
- c) **Queries:** UV disinfection or UV + ozone disinfection or ozone disinfection or UV, UV + COLD D or COLD D disinfection may be asked.
Amendment: No Change

2. Item Name: Cyto Centrifuge

- a) **Queries:** The rpm level may be minimised to 2000rpm.
Amendment: The rpm level should be upto to 2000rpm or higher.
- b) **Queries:** Product certification like USFDA and European CE may be asked.
Amendment: No Change

3. Item Name: Sickle cell Thalassemia and Haemoglobinopathy testing

- a) **Queries:** The item may not have 100 installations in India for all the available company.
Amendment: The quoted model should have atleast one installation in India supplied to Govt./corporate institutions or hospitals.
- b) **Queries:** The HPLC system may not have at least 10 years of presence in India with availability of reagents for thalassemia and Hemoglobinopathy testing.
Amendment: The HPLC system should have at least 5 years of presence in India with availability of reagents for thalassemia and Hemoglobinopathy testing.
- c) **Queries:** The HPLC system may not have a dual piston pump so that each elution buffer has a different pump and the buffers work efficiently.
Amendment: Deleted.

Lot 2/ Package 9: DEPARTMENT OF PHARMACOLOGY

1. Item Name: Digital Plethysmograph

Queries: The technical specification is complete in nature and may be amended.

Amendment: The detail specification is as follows,

- The system should have all in one compact controller unit with Integrated, automated, software regulated bias flow to provide easy plug and play operation.

- The 1 or 2 high quality plethysmographs chamber should be featured with pneumotachograph which should improve the signal-to-noise ratio and reduces unnecessary breath rejection for analysis.
- The system should have a Pre-calibrated digital Temperature and Humidity sensors.
- The Calibration process for the system should be 100% automated, No manual Calibration.
- The unit should have Integrated nebulizer controller and Low BIAS flow noise –minimize signal to noise ratio, critical in WBP application.
- The unit should have software controlled aerosol nebulizer with flow rate 0.1ml/min and Particle size of 5 micron or less and capable of nebulizing in chambers.
- The system should have integrated gas automation facility.
- The complete experiment should be controlled through the software and the hardware.
- The software should provide automated reporting capabilities like Instant reports for Dose Response, PC200, Time Course, Antagonist, Toxicity, and more
- The software should be exported to various file format like .txt, Excel, SPSS™, Graph pad prism™ and other statistical package for desired processing and interpretation.
- Compatible Computer with Intel i5 processor, 4GB RAM, 1TB HDD and 21 inch screen, wifi, Original windows 10, licensed MS office should be provided.

2. Item Name: Kymograph with accessories

Queries: The item no. 1, 10, 11 is a combine system. Hence may be combined as a single item.

Amendment: The revised specification after combining the three items i.e. Kymograph with accessories, Dales isolated organ bath, Accessories for organ bath is as follows:

The System should have:-

- A high quality Perspex single unit organ bath with inbuilt water-pump, gas diffuser, digital temperature controller, filling/draining knobs with easy set up for conducting experiments to record the movements/displacements through the force transducer (0-50 gms) along with required accessories.
- At-least 50 ready to use experiments for animal physiology with sample data & real time recording/analysis & easy customisation of the experiments.
- Simulated Experiments/Pre-configured Experiments list should include :- Chick Biventer Cervicis, Mammalian Atria, Mammalian Jejunum, Mammalian Uterus, Stimulated Ileum, Stimulated Rat Vas Deferens, Toad Rectus Abdominis, Unstimulated Ileum, Unstimulated Rat Vas Deferens, Vascular Resistance Vascular Smooth Muscle, Mammalian Blood Pressure -Drug Effects and Frog Heart Experiments :-Temperature Effects, Starling's Law of the Heart, Drug Effects, Conduction Block, Hear Refractory Period, Vagal Escape, Frog Neuromuscular Junction:- Twitch Recruitment – Nerve, Muscle, force, latency, fatigue, Tubocurarine etc, Frog Sciatic Nerve & Skeletal Muscle:- Twitch Recruitment, Effect of Stretch on Contractile Force, Muscle Summation, Muscle Tetanus, Muscle and Nerve Fatigue, Nerve Threshold, Nerve Refractory Period,
- The software should be provided with free experiments and software updates for at least 5 years.
- It should be supplied with signal conditioners & force transducer (0-50g) and PC with Printer.
- It should allow students and educators to access experiment material as well as the recorded data anywhere from outside the lab through internet.
- System should be compatible with WINDOWS or MACINTOSH operating systems.
- Manufacturer should be ISO & CE Certified with all standard certifications.

3. Item Name: Tail flick apparatus/ Analgesiometer

Queries: Heating platform may be made of thick special grade plastic plate instead of aluminium plate.

Amendment: Heating platform should be made of thick special grade plastic plate.

4. Item Name: Water maze (Mories) for rat and mice with video tracking system, Computerised software to record data

Queries: The software features and configuration of computer system may be incorporated and Video camera & Video monitor requirements may be clearly specified.

Amendment: The amended specification is as follows,

- Computer with Windows XP/7 /8 or 10 Operating System Core i3 or upgraded processor type RAM 2GB or more, Hard Disk 1TB or more USB port colour Monitor should be provided.
- The software should cover wide variety of applications: Motor and locomotor activity Anxiety- open field, plus maze, O maze, Black and white box Water maze, Radial Maze, T maze Emergence test Conditioned Place Preference Novel Object Recognition Supplied with Intex/Logitech make camera.
- Light arrangement in the room should be such that during the video recording, there should not be any reflection of light on the maze. Room size should be such that the light arrangement should not create the light reflections on the maze during video recording. Height of the room should be such that, camera fitted on the top can capture the complete maze with proper visibility of animal on the / in the maze.

5. Item Name: Animal grip Strengthmeter

a) **Queries:** Standard cable length may be 4ft.instead of 6ft.

Amendment: Standard cable length should be atleast 4ft.

b) **Queries:** May have LCD or TFT screen.

Amendment: Should have LCD or TFT screen.

6. Item Name: Rotarod Assembly

a) **Queries:** May be made of black acrylic material instead of Perspex amite (lanes)

Amendment: Should be made of black acrylic material.

c) **Queries:** Acceleration modes may be modified.

Amendment: In acceleration mode, speed range should be 10 to 50RPM in steps of 5RPM and acceleration time of 60 to 300mins.in multiple of 60secs.

7. Item Name: Cook's Pole Climbing Apparatus

a) **Queries:** Electrical shock pulse rate may be changed from 400V to 16 to 200V with rate of 0.1mA.

Amendment: Electrical shock pulse rate should be 16 to 200V with rate of 0.1mA.

b) **Queries:** Output available for recording on kymograph or polygraph complete to work on 220V AC has no relevance in operation.

Amendment: The said point is deleted from the specification.

8. Item Name: Sipper Bottle

b) **Queries:** The specification is obsolete and may be changed.

Amendment: Polycarbonate Water Bottle Capacity of 250 ml. Animals Mice, Rat Cap: Silicon bush with S.S. Cap Press fit Style Sipper tube type: S.S. 304 Nozzle

1. The following items along with required quantity are added under the Lot 3/Package 2 (Department of Physiology) -

Sl no	Items	Qty
1	Volume recorders	6
2	Isolated organs, bath for students complete with liver etc	10
3	Anaesthesia boxes	1
4	Surgical instruments for operative procedures	20

	Syringes, pulleys, fine scissor, long scissor, pithing needle etc.	
5	Douglas bag complete	6
6	Basal metabolism apparatus	1
7	Algometer	3
8	Apparatus for passive movement	1
9	Tuning fork 100-256Hz , 512Hz test hearing	40
10	Kymograph Electric motor Driven	30

2. The following items along with required quantity are added under the Lot 3/Package 5 (Department of Pathology) -

All items should be made of hard borosilicate glass

Items	Unit	Qty.
Museum Jars:140 x 215 x 100mm-		40
Museum Jars 225 x 225 x 125mm		40
Museum Jars 220 x 140 x 140mm		40
Beaker	100ml	50
Beaker	250ml	25
Beaker	500ml	25
Beaker	1000ml	5
Measuring Beaker with handle		5
Burette	10 ml 100 divisions	50
Burette stands	22X15"	30
Burette clamp		50
Bottles for reagents(stoppered)	100 ml	100
Bottles for reagents(stoppered)	250m ml	50
Bottles for reagents(stoppered)	500 ml	25
Bottles for reagents(stoppered)	1000 ml	10
Wash bottles	1000 ml	60
Large water bottles with tap	5000 ml	4
Measuring Cylinders	25 ml	10
Measuring Cylinders	100 ml	10
Measuring Cylinders	500 ml	10
Measuring Cylinders	1000 ml	10
Graduated Centrifuge tubes with cap	15 ml	500
Plain Centrifuge tube	15 ml	500
Gloves	7"	1 packet
Hand Protector grip	7"	1 packet
Glass pipette	1 ml	50
Glass pipette	2 ml	50
Glass pipette	5 ml	50
Glass pipette	10ml	50
Pasteur pipette	3 ml	100

Auto pipettes	10-100 micro litre	5
Auto pipettes	20-200 microlitre	5
Auto pipettes	100-1000 microlitre	5
Rubber teats		50
Measuring Scoops		10 packets
Glass slides for microscope		100
Universal micro tipbox		20
Tips for autopipettes	10-100 micro litre	1000
Tips for autopipettes	20-200 microlitre	1000
Tips for autopipettes	100-1000 microlitre	1000
Test tube	15 ml	1000
Test tube	20 ml	1000
Test tube(Thick)	30 ml	500
Test tube racks		50
Test tube holder		60
Test tube brush different sizes		10
Conical flasks	100 ml	50
Conical flasks	250 ml	20
Conical flasks	500 ml	10
Conical flasks	1000 ml	10
Conical flasks	2000 ml	10
Round bottom flask	100 ml	100
Volumetric flask	500 ml	10
Volumetric flask	1000 ml	10

3. The following items along with required quantity are added under the Lot 3/Package 8 (Department of Pharmacology) -

Items	Units	Numbers
Beaker	100ml	50
Beaker	250ml	25
Beaker	500ml	25
Beaker	1000ml	5
Measuring Beaker with handle		5
Burette	10 ml 100 divisions	50
Burette stands	22X15"	30
Burette clamp		50
Bottles for reagents(stoppered)	100 ml	100
Bottles for reagents(stoppered)	250m ml	50
Bottles for reagents(stoppered)	500 ml	25
Bottles for reagents(stoppered)	1000 ml	10
Wash bottles	1000 ml	60

Large water bottles with tap	5000 ml	4
Measuring Cylinders	25 ml	10
Measuring Cylinders	100 ml	10
Measuring Cylinders	500 ml	10
Measuring Cylinders	1000 ml	10
Graduated Centrifuge tubes with cap	15 ml	500
Plain Centrifuge tube	15 ml	500
Gloves	7"	1 packet
Hand Protector grip	7"	1 packet
Glass pipette	1 ml	50
Glass pipette	2 ml	50
Glass pipette	5 ml	50
Glass pipette	10ml	50
Pasteur pipette	3 ml	100
Auto pipettes	10-100 micro litre	5
Auto pipettes	20-200 microlitre	5
Auto pipettes	100-1000 microlitre	5
Rubber teats		50
Measuring Scoops		10 packets
Glass slides for microscope		100
Universal micro tipbox		20
Tips for auto pipettes	10-100 micro litre	1000
Tips for auto pipettes	20-200 microlitre	1000
Tips for auto pipettes	100-1000 microlitre	1000
Test tube	15 ml	1000
Test tube	20 ml	1000
Test tube (Thick)	30 ml	500
Test tube racks		50
Test tube holder		60
Test tube brush different sizes		10
Conical flasks	100 ml	50
Conical flasks	250 ml	20
Conical flasks	500 ml	10
Conical flasks	1000 ml	10
Conical flasks	2000 ml	10
Round bottom flask	100 ml	100
Volumetric flask	500 ml	10
Volumetric flask	1000 ml	10

Lot 4/ Package 1: DEPARTMENT OF ANATOMY

The following items are removed from Lot 4/ Package 1: Department of Anatomy and included in Lot 1: Common Items and hence, will be evaluated item-wise –

1. Sl. No. 8: Skeleton (Articulated) – 5 Sets
2. Sl. No. 9: Skeleton (Dis-Articulated) – 25 Sets

3. Commercial Parameter:

Amendment to:

A.) Eligibility Criteria

Sl. No.	Provision in the published Bid Document	Modification now made
1.	Page 5: Point No. 7	
	<p>Financial Capability: The Minimum required average annual turnover in respect of the bidder in the last 3 Financial Years i.e. 2013-14, 2014-15 & 2015-16 or 2014-15, 2015-16 & 2016-17 is as follows –</p> <p>a) Lot 1: Common High Value Items - Rs.10 Crore b) Lot 2: Department-wise High Value Items - Rs.10 Crore c) Lot 3: Department-wise Low Value/ Consumables/ Reagents/ Glassware Items/Chemicals - Rs.5 Crore d) Lot 4: Department-wise Models, Charts & Specimens - Rs.2 Crore</p>	<p>Financial Capability: The Minimum required average annual turnover in respect of the bidder in the last 3 Financial Years i.e. 2014-15, 2015-16 & 2016-17 is as follows –</p> <p>a) Lot 1: Common High Value Items: (i) For all Items except Skeleton (Articulated) & Skeleton (Dis-Articulated) - Rs.5 Crore (ii) For Skeleton (Articulated) & Skeleton (Dis-Articulated) - Rs.2 Crore b) Lot 2: Department-wise High Value Items - Rs.10 Crore c) Lot 3: Department-wise Low Value/ Consumables/ Reagents/ Glassware Items/ Chemicals - Rs.2 Crore d) Lot 4: Department-wise Models, Charts & Specimens - Rs.2 Crore</p>
2.	Page 6: Point No. 8 (viii)	
	Copies of PAN Card, Service Tax Registration Certificate, Up to date VAT Clearance Certificate, etc.	Copies of PAN Card, GST Registration Certificate , etc.

B.) General Conditions of Contract

Sl. No.	Provision in the published Bid Document	Modification now made
1.	Page 10: Point No. 11	
	<p>Documents Establishing the Qualifications of the Bidder: The documentary evidence of the Bidder's qualifications to perform the contract if its bid is accepted shall establish to the Purchaser's satisfaction: (a) that, if a Bidder is the Original Equipment Manufacturer (OEM) of the Goods it offers to supply shall submit the Manufacturer's Offer Form using the form included in Annexure VIII; (b) that, if a Bidder that does not manufacture or produce the Goods it offers to supply shall submit the Manufacturer's Authorization using the form included in Annexure IX to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods;</p>	<p>Documents Establishing the Qualifications of the Bidder: The documentary evidence of the Bidder's qualifications to perform the contract if its bid is accepted shall establish to the Purchaser's satisfaction: (a) that, if a Bidder is the Original Equipment Manufacturer (OEM) of the Goods it offers to supply shall submit the Manufacturer's Offer Form using the form included in Annexure VIII; (b) that, if a Bidder that does not manufacture or produce the Goods it offers to supply shall submit the Manufacturer's Authorization using the form included in Annexure IX to demonstrate that it has been duly authorized by the manufacturer or producer of the Goods to supply these Goods;</p>

	<p>(c) that the Bidder meets each of the qualification criterion specified in Eligibility Criteria.</p> <p>(d) Bids from Joint Ventures are not acceptable</p>	<p>(c) that the Bidder meets each of the qualification criterion specified in Eligibility Criteria;</p> <p>(d) Bids from Joint Ventures are not acceptable;</p> <p>(e) Manufacturer's Offer Form (Annexure VIII)/ Manufacturer's Authorization (Annexure IX) is mandatory for items under Lot 1 and Lot 2.</p>
<p>2. Page 15-16: Point No. 24</p>		
	<p>Responsiveness of Bids:</p> <p>24.1 The Purchaser's determination of a bid's responsiveness is to be based on the contents of the bid itself.</p> <p>24.2 A substantially responsive Bid is one that conforms to all the terms, conditions and specifications of the Bidding Documents without material deviation, reservation or omission. A material deviation, reservation or omission is one that:</p> <p>(a) affects in any substantial way the scope, quality or performance of the Goods and Related Services specified in the Contract; or</p> <p>(b) limits in any substantial way, inconsistent with the Bidding Documents, the Purchaser's rights or the Bidder's obligations under the Contract; or</p> <p>(c) if rectified, would unfairly affect the competitive position of other bidders presenting substantially responsive bids.</p> <p>24.3 Bids from Agents, without proper authorization from the manufacturer shall be treated as non-responsive.</p> <p>24.4 If a bid is not substantially responsive to the Bidding Documents, it shall be rejected by the Purchaser and may not subsequently be made responsive by the Bidder by correction of the material deviation, reservation or omission.</p>	<p>24. Responsiveness of Bids:</p> <p>24.1 The Purchaser's determination of a bid's responsiveness is to be based on the contents of the bid itself.</p> <p>24.2 A substantially responsive Bid is one that conforms to all the terms, conditions and specifications of the Bidding Documents without material deviation, reservation or omission. A material deviation, reservation or omission is one that:</p> <p>(a) affects in any substantial way the scope, quality or performance of the Goods and Related Services specified in the Contract; or</p> <p>(b) limits in any substantial way, inconsistent with the Bidding Documents, the Purchaser's rights or the Bidder's obligations under the Contract; or</p> <p>(c) if rectified, would unfairly affect the competitive position of other bidders presenting substantially responsive bids.</p> <p>24.3 Bids from Agents, without proper authorization from the manufacturer, in case of items under Lot 1 and Lot 2, shall be treated as non-responsive.</p> <p>24.4 If a bid is not substantially responsive to the Bidding Documents, it shall be rejected by the Purchaser and may not subsequently be made responsive by the Bidder by correction of the material deviation, reservation or omission.</p>
<p>3. Page 18: Point No. 28</p>		
	<p>Financial Evaluation; Comparison of Bids:</p> <p>The Purchaser shall compare the evaluated prices of all substantially technically responsive bids for each item in case of Lot 1 and for each Package in case of Lot 2, Lot 3 & Lot 4 to determine the lowest evaluated bid for each item in case of Lot 1 and for each Package in case of Lot 2, Lot 3 & Lot 4. Hence, the bidders are requested to quote for all the items of the individual Package under Lot 2, Lot 3 and Lot 4.</p> <p>Prices shall be considered for evaluation like below –</p> <p>The quoted rate should include excise/ customs duty, transportation, insurance, packing & forwarding or any other incidental</p>	<p>Financial Evaluation; Comparison of Bids:</p> <p>The Purchaser shall compare the evaluated prices of all substantially technically responsive bids for each item in case of Lot 1 and for each Package in case of Lot 2, Lot 3 & Lot 4 to determine the lowest evaluated bid for each item in case of Lot 1 and for each Package in case of Lot 2, Lot 3 & Lot 4. Hence, the bidders are requested to quote for all the items of the individual Package under Lot 2, Lot 3 and Lot 4.</p> <p>Prices shall be considered for evaluation like below –</p> <p>The quoted rate should include customs duty, transportation, insurance, packing & forwarding or any other incidental charges.</p>

	<p>charges.</p> <p>28.1 In case of bidders who have quoted CST (firms not registered under Odisha VAT), CST as mentioned in the Price Bid by the bidder shall be added to the quoted rate for price evaluation.</p> <p>28.2 In case of bidders who have quoted VAT (firms registered under Odisha VAT), VAT as mentioned in the Price Bid by the bidder shall be excluded for price evaluation.</p> <p>28.3 Entry Tax will not be considered for price evaluation.</p> <p>28.4 After giving price preferences to eligible local MSE Units of Odisha.</p> <p>28.5 As per the Govt. of Odisha Finance Deptt. Order No.13290/F dt.02.04.2013, “in comparing the cost of an article, if purchased from within the State with the price of similar article if purchased from outside the State, the amount of Odisha Sales Tax (OST), now VAT, shall be deducted from the total cost since it accrues back as revenue to the State. If after such deduction, the cost of articles to be purchased within the State is not more than the cost of including Central Sales Tax, transport and other charges of similar articles from outside the State, it would be economical to purchase articles within the State”.</p> <p>28.6 The basic price, Installation cost (if any), CMC (wherever applicable as mentioned in list of item requirement) and Cost of reagents (wherever applicable) shall be taken into account for evaluation.</p> <p>28.7 Bids which do not contain prices for CMC, wherever applicable, shall be rejected.</p>	<p>28.1 Deleted.</p> <p>28.2 Deleted.</p> <p>28.3 Entry Tax will not be considered for price evaluation.</p> <p>28.4 Deleted.</p> <p>28.5 Deleted.</p> <p>28.6 The basic price, Installation cost (if any), CMC (wherever applicable) and Cost of reagents (In case of close system) shall be taken into account for evaluation.</p> <p>28.7 Bids which do not contain prices for CMC (wherever applicable) and Cost of reagents (In case of close system) shall be rejected.</p>
--	---	---

C.) Special Conditions of Contract

Sl. No.	Provision in the published Bid Document	Modification now made
1.	Page 33: Point No. 20.2	
	This clause will apply only to variations in VAT/ Octroi, etc. payable in India on the final product which is being supplied and not for the individual components / raw materials which go into the product.	This clause will apply only to variations in GST/ Octroi, etc. payable in India on the final product which is being supplied and not for the individual components/ raw materials which go into the product.

D.) Annexures

Sl. No.	Provision in the published Bid Document	Modification now made																																																																																																																																																																																												
1.	Page 137 - Annexure VI/1																																																																																																																																																																																													
	<p><u>PRICE SCHEDULE FOR EQUIPMENTS</u></p> <p>(To be furnished in Cover “B” - Financial Bid)</p> <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 10px;"> <thead> <tr> <th rowspan="2">Sl. No.</th> <th rowspan="2">Name of the Item</th> <th rowspan="2">No. of Units</th> <th rowspan="2">Unit Price (Inclusive of transportation, insurance, installation charge and other services required to convey the Goods to their final destinations) (Rs.)</th> <th rowspan="2">Total Price (Rs.) 5 (=3x4)</th> <th colspan="2">VAT</th> <th colspan="2">CST</th> <th colspan="2">Any Other Tax (Pl. mention)</th> <th rowspan="2">Total Cost (inclusive of Taxes) (Rs.) 12 (=5+9)</th> </tr> <tr> <th>%age</th> <th>Amount (Rs.)</th> <th>%age</th> <th>Amount (Rs.)</th> <th>%age</th> <th>Amount (Rs.)</th> </tr> </thead> <tbody> <tr><td>1</td><td>2</td><td>3</td><td>4</td><td></td><td>6</td><td>7</td><td>8</td><td>9</td><td>10</td><td>11</td><td></td></tr> <tr><td>1</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>2</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>3</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>4</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>5</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>6</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>7</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> </tbody> </table> <p>Total Cost (Col. 12 above): Rs..... (in words) only</p>	Sl. No.	Name of the Item	No. of Units	Unit Price (Inclusive of transportation, insurance, installation charge and other services required to convey the Goods to their final destinations) (Rs.)	Total Price (Rs.) 5 (=3x4)	VAT		CST		Any Other Tax (Pl. mention)		Total Cost (inclusive of Taxes) (Rs.) 12 (=5+9)	%age	Amount (Rs.)	%age	Amount (Rs.)	%age	Amount (Rs.)	1	2	3	4		6	7	8	9	10	11		1												2												3												4												5												6												7												<p><u>PRICE SCHEDULE FOR EQUIPMENTS</u></p> <p>(To be furnished in Cover “B” - Financial Bid)</p> <table border="1" style="width:100%; border-collapse: collapse; margin-bottom: 10px;"> <thead> <tr> <th rowspan="2">Sl. No.</th> <th rowspan="2">Name of the Item</th> <th rowspan="2">No. of Units</th> <th rowspan="2">Unit Price (Inclusive of customs duty, transportation, insurance, packing & forwarding, installation charge and other services required to convey the Goods to their final destinations) (Rs.)</th> <th rowspan="2">Total Price (Rs.) 5 (=3x4)</th> <th colspan="2">GST</th> <th rowspan="2">Total Cost (inclusive of Taxes) (Rs.) 8 (=5+7)</th> </tr> <tr> <th>%age</th> <th>Amount (Rs.)</th> </tr> </thead> <tbody> <tr><td>1</td><td>2</td><td>3</td><td>4</td><td></td><td>6</td><td>7</td><td></td></tr> <tr><td>1</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>2</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>3</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>4</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>5</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>6</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> <tr><td>7</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr> </tbody> </table> <p>Total Cost (Col. 8 above): Rs..... (in words) only</p>	Sl. No.	Name of the Item	No. of Units	Unit Price (Inclusive of customs duty, transportation, insurance, packing & forwarding, installation charge and other services required to convey the Goods to their final destinations) (Rs.)	Total Price (Rs.) 5 (=3x4)	GST		Total Cost (inclusive of Taxes) (Rs.) 8 (=5+7)	%age	Amount (Rs.)	1	2	3	4		6	7		1								2								3								4								5								6								7							
Sl. No.	Name of the Item						No. of Units	Unit Price (Inclusive of transportation, insurance, installation charge and other services required to convey the Goods to their final destinations) (Rs.)	Total Price (Rs.) 5 (=3x4)	VAT		CST		Any Other Tax (Pl. mention)		Total Cost (inclusive of Taxes) (Rs.) 12 (=5+9)																																																																																																																																																																														
		%age	Amount (Rs.)	%age	Amount (Rs.)	%age				Amount (Rs.)																																																																																																																																																																																				
1	2	3	4		6	7	8	9	10	11																																																																																																																																																																																				
1																																																																																																																																																																																														
2																																																																																																																																																																																														
3																																																																																																																																																																																														
4																																																																																																																																																																																														
5																																																																																																																																																																																														
6																																																																																																																																																																																														
7																																																																																																																																																																																														
Sl. No.	Name of the Item	No. of Units	Unit Price (Inclusive of customs duty, transportation, insurance, packing & forwarding, installation charge and other services required to convey the Goods to their final destinations) (Rs.)	Total Price (Rs.) 5 (=3x4)	GST		Total Cost (inclusive of Taxes) (Rs.) 8 (=5+7)																																																																																																																																																																																							
					%age	Amount (Rs.)																																																																																																																																																																																								
1	2	3	4		6	7																																																																																																																																																																																								
1																																																																																																																																																																																														
2																																																																																																																																																																																														
3																																																																																																																																																																																														
4																																																																																																																																																																																														
5																																																																																																																																																																																														
6																																																																																																																																																																																														
7																																																																																																																																																																																														
2.	Page 139 - Annexure VI/3																																																																																																																																																																																													
	Tax (%) : _____. Pl, mention whether CST or OVAT : _____	Tax (%) : _____																																																																																																																																																																																												

Note: The amendments mentioned above are to be treated as amendments in the terms and conditions of the bid document. All other terms & conditions of the published bid document will remain same.

The bidders have to sign and submit copies of this Amendment along with the bid documents. Each page of the bid document as well as this Amendment is to be signed by the authorized Signatories of the bidders.

Sd/-

**Dean & Principal
Govt. Medical College, Balasore**